

CERTIFICATE OF CALIBRATION

ISSUED BY H & D FITZGERALD LTD.

DATE OF ISSUE **31 March 2020**

CERTIFICATE NUMBER **10123**



Cefn Du, Tremeirchion,
St. Asaph, Denbighshire
LL17 0US, UK
☎ +44 (0)1352 720 774
calibration@density.co.uk
www.density.co.uk

Page 1 of 1 pages

Approved signatory

H. Fitzgerald
I. Edge

Description	Dodecane	Grade	99+% purity
Supplier	H&D	Liquid expiry date	30 March 2025
Date calibrated	22 - 29 March 2020		

The density of this liquid was determined by hydrostatic weighing, using instruments and weights, the calibrations of which are traceable to national standards. The liquid in the ampoule formed part of the bulk liquid tested in the hydrostatic weighing cell.

A total of **25** density measurements were made at ambient pressure, and the following densities then calculated:-

10.0°C and 1.013 bar	756.141 kg/m ³ ± 0.010 kg/m ³
15.0°C and 1.013 bar	752.519 kg/m ³ ± 0.010 kg/m ³
20.0°C and 1.013 bar	748.895 kg/m ³ ± 0.010 kg/m ³
25.0°C and 1.013 bar	745.269 kg/m ³ ± 0.010 kg/m ³
30.0°C and 1.013 bar	741.638 kg/m ³ ± 0.010 kg/m ³
40.0°C and 1.013 bar	734.359 kg/m ³ ± 0.010 kg/m ³
50.0°C and 1.013 bar	727.049 kg/m ³ ± 0.010 kg/m ³

The density of this liquid can be found to ± **0.010** kg/m³ over the temperature range of **10** to **50°C** using the following equation:-

$$\text{density} = \alpha + \beta t + \gamma t^2 + \delta t^3$$

where

α	=	763.3840
β	=	- 0.724 577
γ	=	+ 0.000 0398
δ	=	- 0.000 001 65
t		temperature in °C ITS-90

This standard is for the calibration of density measuring equipment.
See our company website for guidance notes on the use of liquid density standards.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of about 95%. The evaluation of the uncertainty has been carried out in accordance with UKAS requirements.

This certificate has been drawn up according to the guidelines given in ILAC-G12:2000.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.